



**Coronavirus  
Disease 2019  
(COVID-19)**

# Fact Sheet for Healthcare Providers

This Fact Sheet informs you of the significant known and potential risks and benefits of the emergency use of the DxTernity SARS-CoV-2 RT-PCR CE Test.

The DxTernity SARS-CoV-2 RT-PCR CE Test is authorized for use with saliva specimens collected from individuals consistent with the EUA.

**All patients whose specimens are tested with this assay will receive the Fact Sheet for Patients: DxTernity Diagnostics, Inc. - DxTernity SARS-CoV-2 RT-PCR CE Test.**

## What are the symptoms of COVID-19?

Many patients with confirmed COVID-19 have developed fever and/or symptoms of acute respiratory illness (e.g., cough, dyspnea). The current information available to characterize the spectrum of clinical illness associated with COVID-19 suggests that symptoms include cough, shortness of breath or dyspnea, fever, chills, myalgias, headache, sore throat or new loss of taste or smell. Based on what is known about the virus that causes COVID-19, signs and symptoms may appear any time from 2 to 14 days after exposure to the virus, and the median incubation period is approximately 5 days. For further information on the symptoms of COVID-19 please see the link provided in *"Where can I go for updates and more information?"* section at the end of this document.

Public health officials have identified cases of COVID-19 infection throughout the world, including the United States. Please check the CDC COVID-19 webpage (see link provided in *"Where can I go for updates and more information?"* section) for the most up to date information.

## What do I need to know about COVID-19 testing?

Current information on COVID-19 for healthcare providers is available at CDC's webpage, Information for Healthcare Professionals (see links provided in *"Where can I go for updates and more information?"* section).

**This test is to be performed only using saliva specimens collected from individuals consistent with the EUA.**

- The DxTernity SARS-CoV-2 RT-PCR CE Test is for use with saliva specimens collected from individuals suspected of COVID-19 by their healthcare provider (HCP)
- The DxTernity SARS-CoV-2 RT-PCR CE Test is also authorized for use with saliva specimens collected from any individual, including individuals without symptoms or other reasons to suspect COVID-19 infection.
- The DxTernity SARS-CoV-2 RT-PCR CE Test is authorized for use on individual saliva specimens that are self-collected at home by individuals using the Spectrum Solutions LLC SDNA-1000 Saliva Collection Device when determined to be appropriate by a healthcare provider based on results of a COVID-19 questionnaire when directly ordered by a HCP.
- The DxTernity SARS-CoV-2 RT-PCR CE Test is authorized for use at DxTernity Diagnostics, Inc., located at 19500 S. Rancho Way, Suite 116, Rancho Dominguez, CA 90220 which is certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, and meets requirements to perform high complexity tests.

Specimens should be collected with appropriate infection control precautions. Current guidance is available at the CDC's website (see links provided in *"Where can I go for updates and more information?"* section).

When collecting and handling specimens from individuals suspected of being infected with COVID-19, appropriate personal protective equipment should be used as outlined in the CDC *Interim Laboratory Biosafety Guidelines for Handling and Processing Specimens Associated with Coronavirus Disease 2019*.

(COVID-19). For additional information, refer to CDC *Interim Guidelines for Collecting, Handling, and Testing Clinical Specimens from Persons Under Investigation (PUIs) for Coronavirus Disease 2019 (COVID-19)* (see links provided in “Where can I go for updates and more information?” section).

### **What does it mean if the specimen tests positive for the virus that causes COVID-19?**

A positive test result for COVID-19 indicates that RNA from SARS-CoV-2 was detected, and therefore the patient is infected with the virus and presumed to be contagious. Laboratory test results should always be considered in the context of clinical observations and epidemiological data in making a final diagnosis and patient management decisions. Patient management should be made by a healthcare provider and follow current CDC guidelines.

The DxTernity SARS-CoV-2 RT-PCR CE Test has been designed to minimize the likelihood of false positive test results. However, in the event of a false positive result, risks to patients could include the following: a recommendation for isolation of the patient, monitoring of household or other close contacts for symptoms, patient isolation that might limit contact with family or friends and may increase contact with other individuals with COVID-19, limits in the ability to work, the delayed diagnosis and treatment for the true infection causing the symptoms, unnecessary prescription of a treatment or therapy, or other unintended adverse effects.

All laboratories using this test must follow the standard testing and reporting guidelines according to their appropriate public health authorities.

### **What does it mean if the specimen tests negative for the virus that causes COVID-19?**

A negative test result for this test means that SARS-CoV-2 RNA was not present in the specimen above the limit of detection. However, a negative result does not rule out COVID-19 and should not be used as the sole basis for treatment or patient management decisions. It is possible to test a person too early or too late during COVID-19 infection to make an accurate diagnosis via the DxTernity SARS-CoV-2 RT-PCR CE Test. In addition, asymptomatic people infected with COVID-19 may not shed enough virus to reach the limit of detection of the test, giving a false negative result. In the absence of symptoms, it is difficult to determine if asymptomatic people have been tested too late or too early. Therefore, negative results in asymptomatic individuals may include individuals who were tested too early and may become positive later, individuals who were

tested too late and may have serological evidence of infection, or individuals who were never infected.

When diagnostic testing is negative, the possibility of a false negative result should be considered in the context of a patient’s recent exposures and the presence of clinical signs and symptoms consistent with COVID-19. The possibility of a false negative result should especially be considered if the patient’s recent exposures or clinical presentation indicate that COVID-19 is likely, and diagnostic tests for other causes of illness (e.g., other respiratory illness) are negative. If COVID-19 is still suspected based on exposure history together with other clinical findings, re-testing using a new sample with a sensitive method should be considered by healthcare providers in consultation with public health authorities. Additional testing may be helpful to ensure testing was not conducted too early.

If a negative result is obtained with a saliva specimen and COVID-19 is still suspected based on exposure history together with other clinical findings, testing an alternative specimen type should be considered by healthcare providers in consultation with public health authorities.

Risks to a patient of a false negative test result include: delayed or lack of supportive treatment, lack of monitoring of infected individuals and their household or other close contacts for symptoms resulting in increased risk of spread of COVID-19 within the community, or other unintended adverse events.

## What is an EUA?

The United States FDA has made this test available under an emergency access mechanism called an Emergency Use Authorization (EUA). The EUA is supported by the Secretary of Health and Human Service's (HHS's) declaration that circumstances exist to justify the emergency use of in vitro diagnostics (IVDs) for the detection and/or diagnosis of the virus that causes COVID-19.

An IVD made available under an EUA has not undergone the same type of review as an FDA-approved or cleared IVD. FDA may issue an EUA when certain criteria are met, which includes that there are no adequate, approved, available alternatives, and based on the totality of scientific evidence available, it is reasonable to believe that this IVD may be effective in diagnosing COVID-19.

The EUA for this test is in effect for the duration of the COVID-19 declaration justifying emergency use of IVDs, unless terminated or revoked (after which the test may no longer be used).

## What are the approved available alternatives?

There are no approved available alternative tests. FDA has issued EUAs for other tests that can be found at: [www.fda.gov/emergency-preparedness-and-response/mcm-legal-regulatory-and-policy-framework/emergency-use-authorization](http://www.fda.gov/emergency-preparedness-and-response/mcm-legal-regulatory-and-policy-framework/emergency-use-authorization)

**Report Adverse events**, including problems with test performance or results, to MedWatch by submitting the online FDA Form 3500 [www.accessdata.fda.gov/scripts/medwatch/index.cfm?action=reporting.home](http://www.accessdata.fda.gov/scripts/medwatch/index.cfm?action=reporting.home)

Or by calling **1-800-FDA-1088**

## Where can I go for updates and more information?

### CDC Webpages:

**General:** [www.cdc.gov/COVID19](http://www.cdc.gov/COVID19)

**Symptoms:** [www.cdc.gov/coronavirus/2019-ncov/symptoms-testing/symptoms](http://www.cdc.gov/coronavirus/2019-ncov/symptoms-testing/symptoms)

**Healthcare Professionals:** [www.cdc.gov/coronavirus/2019-nCoV/guidance-hcp](http://www.cdc.gov/coronavirus/2019-nCoV/guidance-hcp)

**Information for Laboratories:** [www.cdc.gov/coronavirus/2019-nCoV/guidance-laboratories](http://www.cdc.gov/coronavirus/2019-nCoV/guidance-laboratories)

**Laboratory Biosafety:** [www.cdc.gov/coronavirus/2019-nCoV/lab-biosafety-guidelines](http://www.cdc.gov/coronavirus/2019-nCoV/lab-biosafety-guidelines)

**Isolation Precautions in Healthcare Settings:** [www.cdc.gov/coronavirus/2019-ncov/infection-control/control-recommendations](http://www.cdc.gov/coronavirus/2019-ncov/infection-control/control-recommendations)

**Specimen Collection:** [www.cdc.gov/coronavirus/2019-nCoV/guidelines-clinical-specimens](http://www.cdc.gov/coronavirus/2019-nCoV/guidelines-clinical-specimens)

**Infection Control:** [www.cdc.gov/coronavirus/2019-ncov/infection-control/index](http://www.cdc.gov/coronavirus/2019-ncov/infection-control/index)

### FDA Webpages:

**General:** [www.fda.gov/novelcoronavirus](http://www.fda.gov/novelcoronavirus)

**EUAs:** (includes links to patient fact sheet and manufacturer's instructions) [www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/vitro-diagnostics-euas](http://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/vitro-diagnostics-euas)

### DXTertiary Diagnostics, Inc.:

DxTertiary Diagnostics, Inc.  
19500 S. Rancho Way, Suite 116  
Rancho Dominguez, CA 90220

**Customer/Technical Support**  
310-537-7857